



Paper Accepted¹

ISSN Online 2406-0895

Article for practitioners / Rad za praksu

Vedrana Karan^{1,2}, Đula Đilvesi^{2,3}, Mladen Karan^{2,3}, Vladimir Papić^{2,3}, Petar Vuleković^{2,3}

Use of intraoperative neurophysiological monitoring in surgical treatment of spinal diseases

Примена интраоперативног неурофизиолошког мониторинга у оперативном лечењу обољења кичменог стуба

¹University of Novi Sad, Faculty of Medicine, Department of Physiology, Novi Sad, Serbia;

²University of Novi Sad, Faculty of Medicine, Department of Surgery, Novi Sad, Serbia;

³Clinical Center of Vojvodina, Clinic of Neurosurgery, Novi Sad, Serbia

Received: July 9, 2018

Accepted: November 23, 2018

Online First: December 31, 2018

DOI: <https://doi.org/10.2298/SARH180709076K>

¹**Accepted papers** are articles in press that have gone through due peer review process and have been accepted for publication by the Editorial Board of the *Serbian Archives of Medicine*. They have not yet been copy edited and/or formatted in the publication house style, and the text may be changed before the final publication.

Although accepted papers do not yet have all the accompanying bibliographic details available, they can already be cited using the year of online publication and the DOI, as follows: the author's last name and initial of the first name, article title, journal title, online first publication month and year, and the DOI; e.g.: Petrović P, Jovanović J. The title of the article. *Srp Arh Celok Lek*. Online First, February 2017.

When the final article is assigned to volumes/issues of the journal, the Article in Press version will be removed and the final version will appear in the associated published volumes/issues of the journal. The date the article was made available online first will be carried over.

²**Correspondence to:**

Vedrana KARAN

Faculty of Medicine, Department of Physiology, Hajduk Veljkova 3, 21000 Novi Sad

Email: vedrana.karan@mf.uns.ac.rs

Use of intraoperative neurophysiological monitoring in surgical treatment of spinal diseases

Примена интраоперативног неурофизиолошког мониторинга у оперативном лечењу обољења кичменог стуба

SUMMARY

Intraoperative neurophysiological monitoring (IONM) is a process which uses different kinds of electrophysiological methods with the purpose to monitor and map neural structures during surgical procedures with the aim to preserve them. The fact that we know how important the function of the spine, spinal cord, nerve roots and every other structure is, it is obvious how significant it is to preserve all of them in surgical procedures where these structures are under risk. Although the frequency of neurological complications is not high, it is considered that these deficits are devastating complications of spinal surgery which can have serious consequences on the quality of life and increase health care costs. Because of that, we emphasize the methods of intraoperative neurophysiological development, which provide optimal monitoring of the spinal cord function during routine and complex spinal procedures and has high efficacy detecting possible neurological deficits. The concept of multimodal neuromonitoring, which is used today, relies on advantages of each modality separately and then in combination with these modalities it achieves a more reliable estimation of functional integrity. Today IONM is routinely used worldwide, but in Serbia, its use is still limited even though its advantages are well known.

Keywords: intraoperative neuromonitoring; neurological complications, spine surgery

САЖЕТАК

Интраоперативни неурофизиолошки мониторинг (ИОНМ) подразумева употребу различитих електрофизиолошких метода у сврху праћења функционалног интегритета и мапирања одговарајућих нервних структура током хируршке интервенције са циљем њиховог очувања. Ако знамо колико је важна улога кичме, кичмене мождине, нервних коренова, јасно је колико је важно очувати их током операција у којима постоји могућност њиховог оштећења. Иако стопа неуролошких компликација није висока, оштећења ових структура се сматрају поражавајућом компликацијом спиналне хирургије и могу имати значајан утицај на квалитет живота и повећање трошкова лечења ових пацијената. Због тога се акценат ставља на развој метода ИОНМ које омогућавају оптимално праћење функције кичмене мождине и нервних коренова током рутинских и комплексних захвата у спиналној хирургији, и имају високу ефикасност у детекцији могућих неуролошких оштећења. Концепт мултимодалног неуромониторинга који се данас користи ослања се на предност сваког модалитета посебно, а њиховом комбинацијом постиже се много поузданија процена функционалног интегритета. Данас се ИОНМ користи рутински широм света, међутим код нас је његова употреба још увек ограничена, иако су њене предности добро познате.

Кључне речи: интраоперативни неуромониторинг; неуролошке компликације; операција кичме

Intraoperative neurophysiological monitoring (IONM) is a method which uses different kinds of electrophysiological methods for monitoring and mapping of neural structures during surgical procedures with aim to preserve them. Certain kinds of IONM were used in the early 1970s and 1980s. Their use became everyday practice in different types of surgery, especially in neurosurgery and orthopedic surgery. Today IONM is routinely used worldwide, but in Serbia its use is still limited, even though the advantages are well known.

Pathology of the spine is very diverse. There are deformities, degenerative diseases, injuries and tumors which could be primary and metastatic. The conservative therapy can be used in some occasions, but surgery is often the only possible option. If we know how important the function of the spine, spinal cord, nerve roots and all other structures are, it is obvious how important it is to preserve them in surgical procedures where these structures are under risk. Recent data from literature report that new deficits in scoliosis surgery in adolescent population vary between 0.4–4.5% and in adult population 1.9–2.4% [1, 2]. Cramer et al. [3] reported the rate of neurological deficits of 0.178%, in a ten year-long retrospective study which included degenerative, neoplastic, traumatic and infectious disease. Hamilton et al. [4] reported new neurological deficit in 1% of cases in their study from 2011 which included 108419 all type of spine procedures, in adult and pediatric population. Although the rate of neurological complications is not so high it is considered that these deficits are devastating complications of spinal surgery. Even if complications are rare, there is always a risk of neural structures injuries which can have serious consequences on the quality of life and an increase of health care costs [5, 6, 7]. Because of that the accent is on the developing of intraoperative neurophysiological methods which could reliably monitor spinal cord during the surgical interventions.

Depending on spine pathology, symptoms could be various, but neurological deficits of varying degrees are very often present. The most often pathologies are certainly degenerative diseases of spine. The first symptom includes pain, usually lumbar pain, neck pain or rarely arm pain, and thoracic pain is the rarest. Disc herniation in the cervical spine is usually at C5-6 level and C6-7 level, and anterior cervical discectomy and fusion is one of the most frequently performed procedures in neurosurgery. Depending on the herniated disk position and affected neural structures, clinical appearance can be either in the form of radiculopathy, myelopathy or both. Different authors reported complications in treatment of

cervical myelopathy, with the most severe consequence being cervical medulla compression, from 4.4–20% [8, 9]. The rate in patients without myelopathy complications is very low 0.09–0.6% [10, 11]. Kelleher et al. [12] found that sensitivity of SSEP was 52% and specificity 100% in a study which included 1055 cases with operation on the cervical spine. Sensitivity of MEP was 100% and specificity 96%, sensitivity of EMG was 46% and specificity 73%. In the lumbar spine the most commonly affected levels are L4-5 and L5-S1. Gunnarson et al. [13] found sensitivity of SSEP 28.6% and specificity of 98.7%, EMG sensitivity was 100% and specificity 23.7% in the lumbar spine procedures. Therefore, it is considered that combined use of different modalities of IONM in these procedures provide higher sensitivity and specificity, allow timely intervention, decrease postoperative complications, and improve final outcome [14].

Primary tumors of the spine are rare, but metastatic tumors are common. Tumors can affect bone structures of the spine or they can be intraspinal. There are a few subgroups of intraspinal tumors. They can be extradural and intradural, while intradural tumors can be extramedullary and intramedullary. These tumors could be benign or malign, but due to their localization they have a very high risk of neural complications during surgery, particularly intramedullary tumors.

During the surgical procedures neurological deficits can arise from direct surgical injury of neural tissue, compression, traction, or compromised blood supply in neural tissue [15]. Compressive spinal epidural or subdural hematoma can occur after surgical treatment and can be detected early using IONM [16]. In degenerative spinal disease mechanical compression from ligamentum flavum, posterior longitudinal ligament, intervertebral disc or bone structures can affect neural elements [16]. Positioning is also a crucial factor which can cause compression or neural structures injury during spinal procedures [17]. The use of

IONM in spinal surgery significantly decreases the risk of intraoperative damage of neural structures during each phase of a surgery and improves the outcome of surgical treatment.

Somatosensory evoked potentials (SSEP) were primarily used in scoliosis surgery and they significantly improved positive outcome of these operations. It is considered that the use of SSEP reduces paraplegia for 60%. SSEP estimate dorsal column medial lemniscus system from the periphery to the somatosensory cortex. Stimulation is performed on periphery mixed nerves, for upper extremities typically on median or ulnar nerve, while for lower extremities on posterior tibial nerve or common peroneal nerve. Needle and surface electrodes are usually used. Registration can be performed at appropriate places on the scalp, above somatosensory cortex, according to 10–20 International Electrode System, but also on subcortical and peripheral level. At the intraoperative phase of procedures, before and after positioning, we set the base line up, which serves as a control for potentials obtained during the operation. They can be recorded continuously throughout the operation, while they change the amplitude and latency, which are compared to the base line. At that moment, it is necessary to analyze all the variations in the context of technical issues, anesthesia effects, general effects and surgical manipulation. Alarm criteria is the amplitude decrease for 50% with or without the increase of latency [18]. It is very important to keep in mind sensitivity to inhalational anesthetics, systemic factors such as hypothermia, hypotension and hypoperfusion. Also, all decreases of amplitude are not clinically significant. If amplitude decreases on 50% or lower of the base line, risk is higher, but an appropriate and timely reaction can lead to SSEP recovery and it can preserve the function. It is crucial whether changes occur gradually or suddenly. The most difficult is a sudden and complete loss of potentials without recovery. Stable intraoperative potentials are good predictors of a positive neurological outcome. Nuwer et al. [19] estimated the clinical efficacy of SSEP monitoring

during scoliosis surgery in their extensive multicentric study, and they found that sensitivity was 92% and specificity 98.9% in detection of postoperative neurological deficits.

Motor evoked potentials (MEP) managed to take standard place in IONM in the last three decades. In the 1980s, it was shown that transcranial use of high voltage pulses can induce contralateral motor activity. This technique is very sensitive to anesthesia, so during the time multi pulse stimulation technique was developed with variation of anesthetics use. In the 1990s, this technique became a routine for monitoring corticospinal tract. These potentials are safe and reliable for use in spinal procedures [20]. The most common way for eliciting MEP is transcranial electrical stimulation with electrodes placed on the scalp over the motor cortex according to 10–20 International Electrode System and also with direct cortical stimulation. The most suitable for use on the scalp are corkscrew electrodes because of their stability and low impedance, and for direct cortical stimulation strip electrodes are used. Registration can be done on peripheral muscles, on compound muscle action potential (CMAP), which are the result of α motor neuron activation and on spinal cord as the D wave. Registration from muscles is performed with needle electrodes which are placed on the appropriate places. D wave registration is performed with special electrodes which can be placed in epidural or subdural space. MEP are very sensitive to anesthetics and myorelaxants. Anesthesia is based on the use of propofol and opioids such as fentanyl and remifentanyl, while myorelaxants can be given only at the beginning for intubation. It is very important that D wave is not under the effects of anesthesia. Amplitude and latency of D wave vary depending on the spinal cord level. If the change is more caudal, amplitude will decrease and latency will increase. Below the T9 level it is very difficult to record a reliable D wave. It is considered that D wave is the most important during intramedullary spinal tumor operations [21]. At the beginning of the surgery, we set up the base line before and after positioning of a patient. Amplitude, latency and intensity of stimulation are monitored. Also, all changes of

MEP are considered in the context of anesthesia, systemic effects, surgical manipulation and their development, whether they be gradual or sudden. There are different criteria which can warn us that MEP changes could be significant. One group of authors suggest intensity of stimulation, others propose changes of amplitude, while some recommend only presence or absence of MEP [22, 23, 24]. Quinones-Hinojosa et al. [25] state that changes in amplitude and reduction of complexity of MEP curve correlate with motor outcome. Multichannel monitoring of MEP has higher specificity, sensitivity and prediction of postoperative motor deficits [26]. In spine surgery of intramedullary tumors, one of the most reliable criteria is combination of MEP and D wave. Complete loss of MEP without changes of D wave, or with changes above 50% of D wave amplitude correlate with temporary motor deficit. Complete loss of MEP and decrease below 50% of D wave amplitude, or loss of D wave is a predictor of permanent deficit [27].

Electromyography (EMG) records electrical activity of muscles. It can be a free-run EMG, which registers spontaneous muscle activity and it allows continuous monitoring, and it can be triggered EMG, which implies direct stimulation of peripheral motor nerves or spine roots and registration of CMAP in the appropriate muscle. Surgical manipulations in form of traction, dragging and compression lead to activation of specific muscles and that could be registered on free run EMG. Changes can be in the form of spikes, bursts and trains. Spikes and bursts give us information about the vicinity of a nerve root and they usually appear because of the contact with surgical instruments [28]. Trains appear when continuous force acts on nerve roots, and are clinically significant because they indicate possible injuries [28]. Use of triggered EMG is highly recommended for adequate positioning of pedicle screws, because breach of pedicle cortex can cause injuries of nerve roots and spinal cord. In anesthesia, paralytic agents are contraindicated, with the exception of myorelaxants when anesthesia is first introduced. Relaxation is checked with standard train of four method. Also,

it is very important to determine whether there are comorbidities such as myasthenia gravis, muscle dystrophy or similar pathology, which can have influence on EMG.

The concept of multimodal monitoring used today relies on advantages of each modality separately and then, in combination of these modalities, it achieves more reliable functional integrity estimation. In the study which included spinal procedures of deformities, spinal stenosis and spinal tumors, Sutter et al. [29] found sensitivity of multimodal monitoring of 89% and specificity of 99%. American Academy of Neurology and American Clinical Neurophysiology Society guidelines recommend the use of intraoperative monitoring of MEP and SSEP in spinal surgery as an effective tool in prediction of increased risk of neurological complications [30].

The combination of SSEP, MEP, free-run and triggered EMG provides optimal monitoring of the spinal cord function during routine and complex spinal procedures and has high efficacy in detecting of possible neurological deficits.

REFERENCES

1. Qiu Y, Wang S, Wang B. Incidence and risk factors of neurological deficits of surgical correction for scoliosis; analysis of 1373 cases at one Chinese institution. *Spine*. 2008; 33(5):519–26. [doi: 10.1097/BRS.0b013e3181657d93]
2. Smith JS, Shaffrey CI, Glassman SD. Risk-benefit assessment of surgery for adult scoliosis: an analysis based on patient age. *Spine*. 2011; 36(10):817–24. [doi: 10.1097/BRS.0b013e3181e21783]
3. Cramer DE, Maher PC, Pettigrew DB, Kuntz C. Major neurologic deficit immediately after adult spinal surgery: incidence and etiology over 10 years at a single training institution. *J Spinal Disord Tech*. 2009; 22:565–70. [doi: 10.1097/BSD.0b013e318193452a]
4. Hamilton DK, Smith JS, Sansur CA, Glassman SD, Ames CP, Berven SH, et al. Rates of New Neurological Deficit Associated With Spine Surgery Based on 108,419 Procedures. *Spine*. 2011; 36(15):1218–28. [doi: 10.1097/BRS.0b013e3181ec5fd9]
5. Nasser R, Yadla S, Maltenfort MG, Harrop JS, Anderson DG, Vaccaro AR, et al. Complications in spine surgery. *J Neurosurg Spine*. 2010; 13(2):144–57. [doi: 10.3171/2010.3.SPINE09369]
6. Campbell PG, Yadla S, Malone J, Maltenfort MG, Harrop JS, Sharan AD, et al. Complications related to instrumentation in spine surgery: a prospective analysis. *Neurosurg Focus*. 2011; 31(4):10. [doi: 10.3171/2011.7.FOCUS1134]
7. Ney JP, van der Goes DN, Watanabe JH. Cost-benefit analysis: intraoperative neurophysiological monitoring in spinal surgeries. *J Clin Neurophysiol* 2013; 30:280–6. [doi: 10.1212/WNL.0000000000002076]
8. Khan M, Smith P, Balzer J, Crammond D, Welch W, Gerszten P, et al. Intraoperative somatosensory evoked potential monitorin during cervical spine corpectomy surgery: Experience with 508 cases. *Spine*. 2006; 31(4):E105–13. [doi: 10.1097/01.brs.0000200163.71909.1f]
9. Peolsson A, Peolsson M. Predictive factors for long-term outcome of anterior cervical decompression and fussion: A multivariate data analysis. *Eur Spine J*. 2008; 17:406–14. [doi: 10.1007/s00586-007-0560-2]
10. Ajiboye R, Zoller S, Sharma A, Mosich G, Drysch A, Li J, et al. Intraoperative Neuromonitoring for Anterior Cervical Spine Surgery. *Spine*. 2017; 42(6):385–93. [doi: 10.1097/BRS.0000000000001767]
11. James WS, Rughani AI, Dumont TM. A socioeconomic analysis of IONM during spine surgery: national use, regional variation, and patient outcomes. *Neurosurg Focus*. 2014; 37(5):E10. [doi: 10.3171/2014.8.FOCUS14449]
12. Kelleher MO, Tan G, Sarjeant R, Fehlings MG. Predictive value of intraoperative neurophysiological monitoring during cervical spine surgery: a prospective analysis of 1055 consecutive patients. *J Neurosurg Spine*. 2008; 8(3):215–21. [doi: 10.3171/SPI/2008/8/3/215]
13. Gunnarsson T, Krassioukov AV, Sarjeant R, Fehlings MG. Real-time continuous intraoperative electromyographic and somatosensory evoked potential recordings in spinal surgery: correlation of clinical and electrophysiological findings in a prospective, consecutive series of 213 cases. *Spine*. 2004; 29(6):677–84. [PMID: 15014279]
14. Sutter M, Eggspuehler A, Grob D, Porchet F, Jeszenszky D, Dvorak J. multimodal intraoperative monitoring (MIOM) during 409 lumbosacral surgical procedures in 409 patients. *Eur Spine J*. 2007; 16(2):S221–8. [doi:10.1007/s00586-007-0432-9]
15. Orchowski J, Bridwell KH, Lenke LG. Neurological deficit from a purely vascular etiology after unilateral vessel ligation during anterior thoracolumbar fusion of the spine. *Spine*. 2005; 30(4):406–10. [PMID: 15706337]
16. Sokolowski MJ, Garvey TA, Perl J, Sokolowski MS, Cho W, Mehbod AA, et al. Prospective study of postoperative lumbar epidural hematoma: incidence and risk factors. *Spine*. 2008; 33(1):108–13. [doi: 10.1097/BRS.0b013e31815e39af]

17. Raynor B, Bright JD, Lenke LG, Ra'Kerry K, Bridwell KH, Riew KD, et al. Significant Change or Loss of Intraoperative Monitoring Data: A 25-Year Experience in 12,375 Spinal Surgeries. *Spine*. 2013; 38(2):E101–8. [doi: 10.1097/BRS.0b013e31827aafb9]
18. American Electroencephalographic Society. Guideline eleven: guidelines for intraoperative monitoring of sensory evoked potentials. *J Clin Neurophysiol*. 1994; 11:77–87.
19. Nuwer MR, Dawson EG, Carlson LG, Kanim LE, Sherman JE. Somatosensory evoked potential spinal cord monitoring reduces neurologic deficits after scoliosis surgery: results of a large multicenter survey. *Electroencephalogr Clin Neurophysiol*. 1995; 96(1):6–11. [PMID: 7530190]
20. de Haan P, Kalkman CJ. Spinal cord monitoring: somatosensory- and motor- evoked potentials. *Anesthesiol Clin North America*. 2001; 19(4):923–45. [doi.org/10.1016/S0889-8537(01)80017-1]
21. Deletis V, Sala F. Intraoperative neurophysiological monitoring of the spinal cord during spinal cord and spine surgery: a review focus on the corticospinal tracts. *Clin Neurophysiol*. 2008; 119(2):248–64. [DOI: 10.1016/j.clinph.2007.09.135]
22. Calancie B, Harris W, Brindle GF, Green BA, Landy HJ. Treshold-level repetitive transcranial electrical stimulation for intraoperative monitoring of central motor conduction. *J Neurosurg*. 2001; 95(2):161–8. [PMID:11599831]
23. Hilibrand AS, Schwartz DM, Sethuraman V, Vaccaro AR, Albert TJ. Comparison of transcranial electric motor and somatosensory evoked potential monitoring during cervical spine surgery. *J Bone Joint Surg*. 2004; 86(6):1248–53. [PMID: 15173299]
24. Jones SJ, Harrison R, Koh KF, Mendoza N, Crockard HA. Motor evoked potential monitoring during spinal surgery: responses of distal limb muscles to transcranial cortical stimulation with pulse trains. *Electroencephalogr Clin Neurophysiol*. 1996; 100:375–83. [PMID: 8893655]
25. Quinones-Hinojosa A. Changes in transcranial motor evoked potential during intramedullary spinal cord tumor resection correlate with postoperative motor function. *Neurosurgery*. 2005; 56(5):982–93. [PMID: 15854246]
26. Dong-Gun K, Seong-Rae J, Young-Seop P, Seung-Jae H, Ki-Jeong K, Tae-Ahn J, et al. Multi-channel motor evoked potential monitoring during anterior cervical discectomy and fusion. *Clinical Neurophysiology Practice* 2017; 2:48–53. [doi.org/10.1016/j.cnp.2016.12.006]
27. Deletis V, Sala F. the role of intraoperative neurophysiology in the protection or documentation of surgically induced injury to the spinal cord. *Ann NY Acad Sci*. 2001; 939:137–44. [PMID: 11462765]
28. Gonzalez A, Jeyanandarajan D, Hansen C, Zada G, Hsieh P. Intraoperative neurophysiological monitoring during spine surgery: a review. *Neurosurg Focus*. 2009; 27(4):E6. [doi: 10.3171/2009.8.FOCUS09150]
29. Sutter M, Eggspuehler A, Muller A, Dvorak J. Multimodal intraoperative monitoring: an overview and proposal of methodology based on 1017 cases. *Eur Spine J*. 2007; 16(2):153–61. [doi: 10.1007/s00586-007-0417-8]
30. Nuwer MR, Emerson RG, Galloway G, Legatt AD, Lopez J, Minahan R, et al. Evidence-based guideline update: intraoperative spinal monitoring with somatosensory and transcranial electrical motor evoked potentials: report of the Therapeutics and Technology Assessment Subcommittee of the American Academy of Neurology and the American Clinical Neurophysiology Society. *Neurology*. 2012; 79(3):585–9. [doi: 10.1212/WNL.0b013e3182637c24]